


# Tebuconazole



CIPAC Collaborative Trial  
CIPAC 5374M, full scale study


CIPAC Collaborative Study of a  
Gas Chromatographic Analysis of Tebuconazole Technical Material  
and Formulated Products

By

Mr. Terry Wang

Jiangsu Sevencontinent Green Chemical Co., Ltd.

Presented at the CIPAC meeting in Wageningen  
June, 2024



<i>ISO common name</i>	Tebuconazole
<i>Chemical name</i>	(RS)-1-P-(4-chlorophenyl-4,4-dimethyl-3-(1H-1,2,4-triazole-ylmethyl)pentan-1-ol (IUPAC); (±)α-[2-(4-chlorophenyl)ethyl]- α-(1,1-dimethylethyl)-1H-1,2,4-triazole-1-ethanol (CA)
<i>CAS No</i>	107534-96-3
<i>Empirical formula</i>	C <sub>16</sub> H <sub>22</sub> ClN <sub>3</sub> O
<i>RMM.</i>	307.8
<i>Melting point</i>	99.2±0.5°C (372±0.5K)
<i>Vapor pressure</i>	3.1×10 <sup>-8</sup> mPa (25°C)
<i>Density</i>	1.25 (26°C)
<i>Solubility</i>	In water 32 g/L, (20°C); In Methanol >250g/L, (25°C) In 1,2-dichloroethane >250g/L, (25°C)
<i>Description</i>	White powder.
<i>Stability</i>	Stable to elevated temperatures, and to photolysis and hydrolysis in pure water, under sterile conditions; hydrolysis DT50 >1 y (pH 4-9, 22 °C)
<i>Formulations</i>	Wettable Powder, Emulsifiable Concentrate , Suspension Concentrate, Emulsion in water

**TEBUCONAZOLE TECHNICAL  
494/TC/(M)/-**

**1 Sampling.** Take at least 100 g.

**2 Identity tests**

**2.1 Infrared.** prepare a film between potassium bromide plates and scan from 4000 to 400cm<sup>-1</sup>. The spectrum obtained from the sample should not differ significantly from that of the reference grade material. (Fig. 1)

**2.2 GLC.** Use the GLC method described below. The relative retention time of Tebuconazole in the sample solution should not deviate by more than 1% from that of the calibration solution.

**3 Tebuconazole**

**OUTLINE OF THE METHOD.**

The content of Tebuconazole (g/kg) is determined by capillary gas chromatography with split injection, using dicyclohexyl phthalate as internal standard.

**REAGENTS**

*Tebuconazole* reference standard, of known content

*Dicyclohexyl phthalate* Internal standard. Must not contain impurities with the same retention time as Tebuconazole.

*Acetone* reagent grade ≥99.5%.

**Internal standard solution.** Prepare a single stock of 6mg/ml internal standard solution, of sufficient volume for all samples to be analyzed. For example, to prepare 250ml stock solution, dissolve 1.5g (to the nearest 0.01 g) dicyclohexyl phthalate in acetone in a volumetric flask (250 ml) and make up to the mark with the same solvent.

**Calibration solutions C1 and C2.** Weigh in duplicate (to the nearest 0.01 mg) approximately 50mg (s in mg) of the Tebuconazole reference standard into separate suitable vessels. Add by pipette internal standard solution (10 ml). Mix thoroughly. (calibration solutions C1, C2, chromatogram of C1 see Fig. 2).

**APPARATUS**

*Gas chromatograph* equipped with a split/splitless injection and a flame ionization detector.

*Capillary column*, fused silica, 30 m x 0.32 (i.d.) mm, with a HP-5 bounded phase and a film thickness of 0.25 μm, or equivalent with the same selectivity.

*Electronic integrator or data system*

*Ultrasonic bath*

**PROCEDURE**

**(a) Operating conditions (typical):**

<i>Injector type</i>	split injection
<i>Split Ratio</i>	20:1
<i>Injection volume</i>	0.2 μL
<i>Injector temperature</i>	280 °C
<i>Detector type</i>	flame ionization

Detector temperature	300 °C
Oven temperature	240 °C hold 8min, ramp rate 15°C/min, to 260°C, hold 4 min
<i>Flow rates</i>	carrier gas helium: 2mL/min make-up gas helium: 25 mL/min air 400 mL/min hydrogen 40 mL/min
<i>Running time</i>	14 minutes
<i>Retention time</i>	Tebuconazole: approx. 4.8 min dicyclohexyl phthalate: approx. 6.3min

**(b) System equilibration.** Pump sufficient carrier gas through the column to equilibrate the system. Inject 0.2 µL portions of the calibration solution C1 and repeat the injections until retention times and peak areas deviate by less than ± 1 % from the mean for three successive injections.

**(c) Preparation of sample solution.** Prepare sample solutions in duplicate for each sample. Weigh (to the nearest 0.01 mg) sufficient sample (*w* in mg) (containing approximately 50 mg of Tebuconazole) into separate suitable vessels. Add by pipette internal standard solution (10 ml). Mix thoroughly. (Sample solutions S1, S2, chromatogram of S1 see Fig. 3)

**(d) Determination.** Inject in duplicate each sample solution and bracket a series of sample solution injections by injections of the calibration solution as follows: calibration solution C1, calibration solution C2, calibration solution C1, sample solution S1, sample solution S1, calibration solution C1, sample solution S2, sample solution S2, calibration solution C1 ... (C1, C2, C1, S1, S1, C1, S2, S2, C1 ...)

Determine the peak areas of Tebuconazole and dicyclohexyl phthalate.

**(e) Calculation.** Calculate the mean value of each pair of calibration response factors bracketing the two injections of a sample and use this value for calculating the Tebuconazole contents of the bracketed sample injections.

$$f_i = \frac{I_r \times s \times P}{H_s}$$

$$\text{Content of Tebuconazole} = \frac{f \times H_w}{I_q \times w} \text{ g/kg}$$

where:

$f_i$  = individual response factor

$f$  = mean response factor

$H_s$  = peak area of Tebuconazole in the calibration solution

$H_w$  = peak area of Tebuconazole in the sample solution

$I_r$  = peak area of the internal standard in the calibration solution

$I_q$  = peak area of the internal standard in the sample solutions

s = mass of Tebuconazole reference standard in the calibration solution (mg)

w = mass of sample taken (mg)

P = purity of Tebuconazole reference standard (g/kg)

**TEBUCONAZOLE WETTABLE POWDER  
494/WP/ (M) /-**

**1 Sampling.** Take at least 100 g.

**2 Identity tests.**

**2.1 GLC.** As for tebuconazole 494/TC/M/-

**2.2 GC-MS.** As for tebuconazole 494/TC/M/-

**3 tebuconazole.**

Same approach as for tebuconazole 494/TC/M/-

**3.1 Determination of tebuconazole by gas chromatography**

As for tebuconazole 494/TC/M/- in addition:

Disposable PTFE syringe filter compatible with organic solvents and a 0.45 µm pore diameter or centrifuge.

**PROCEDURE**

*(c) Preparation of sample solution.* Prepare sample solutions in duplicate for each sample. Weigh (to the nearest 0.01 mg) sufficient sample (*w* in mg) (containing approximately 50 mg of tebuconazole) into separate suitable vessels. Add by pipette internal standard solution (10 ml). Mix thoroughly. Clarify a part of the solution by centrifugation or filtration prior to analysis.

(sample solutions S3, S4, chromatogram of S3 see Fig. 4).

**TEBUCONAZOLE SUSPENSION CONCENTRATE  
494/SC/M/-**

**1 Sampling.** Take at least 500 mL. Shake the sample well before weighing.

**2 Identity tests.**

**2.1 GLC.** As for tebuconazole 494/TC/M/-

**2.2 GC-MS.** As for tebuconazole 494/TC/M/-

**3 tebuconazole.**

Same approach as for tebuconazole 494/TC/M/-

### 3.1 Determination of tebuconazole by gas chromatography

As for tebuconazole 494/TC/M/- in addition:

*Calibration solutions C1 and C2.* Weigh in duplicate (to the nearest 0.01 mg) approximately 50 mg (*s* in mg) of the tebuconazole reference standard into separate suitable vessels. Add by pipette internal standard solution (10 ml). Mix thoroughly. (calibration solutions C1, C2, chromatogram of C1 see Fig. 2).

Disposable PTFE syringe filter compatible with organic solvents and a 0.45 µm pore diameter or centrifuge.

#### PROCEDURE

(c) *Preparation of sample solution.* Prepare sample solutions in duplicate for each sample. Weigh (to the nearest 0.01 mg) sufficient sample (*w* in mg) (containing approximately 50 mg of tebuconazole) into separate suitable vessels. Add by pipette internal standard solution (10 ml). Mix thoroughly. (sample solutions S5, S6, chromatogram of S5 see Fig. 5).

## TEBUCONAZOLE EMULSIFIABLE CONCENTRATE 494/WP/M/-

**1 Sampling.** Take at least 500 mL. Shake the sample well before weighing.

### 2 Identity tests.

**2.1 GLC.** As for tebuconazole 494/TC/M/-

**2.2 GC-MS.** As for tebuconazole 494/TC/M/-

### 3 tebuconazole.

Same approach as for tebuconazole 494/TC/M/-

### 3.1 Determination of tebuconazole by gas chromatography

As for tebuconazole 494/TC/M/- in addition:

*Calibration solutions C1 and C2.* Weigh in duplicate (to the nearest 0.01 mg)

approximately 50 mg (*s* in mg) of the tebuconazole reference standard into separate suitable vessels. Add by pipette internal standard solution (10 ml). Mix thoroughly. (calibration solutions C1, C2, chromatogram of C1 see Fig. 2).

#### PROCEDURE

(c) *Preparation of sample solution.* Prepare sample solutions in duplicate for each sample. Weigh (to the nearest 0.01 mg) sufficient sample (*w* in mg) (containing approximately 50 mg of tebuconazole) into separate suitable vessels. Add by pipette internal standard solution (10 ml). Mix thoroughly. (sample solutions S7, S8, chromatogram of S7 see Fig. 6).

### TEBUCONAZOLE EMULSION IN WATER 494/WP/M/-

**1 Sampling.** Take at least 500 mL. Shake the sample well before weighing.

#### **2 Identity tests.**

**2.1 GLC.** As for tebuconazole 494/TC/M/-

**2.2 GC-MS.** As for tebuconazole 494/TC/M/-

#### **3 tebuconazole.**

Same approach as for tebuconazole 494/TC/M/-

#### **3.1 Determination of tebuconazole by gas chromatography**


As for tebuconazole 494/TC/M/- in addition:

*Calibration solutions C1 and C2.* Weigh in duplicate (to the nearest 0.01 mg) approximately 50 mg (*s* in mg) of the tebuconazole reference standard into separate suitable vessels. Add by pipette internal standard solution (10 ml). Mix thoroughly. (calibration solutions C1, C2, chromatogram of C1 see Fig. 2).


#### PROCEDURE




(c) *Preparation of sample solution.* Prepare sample solutions in duplicate for each sample. Weigh (to the nearest 0.01 mg) sufficient sample ( $w$  in mg) (containing approximately 50 mg of tebuconazole) into separate suitable vessels. Add by pipette internal standard solution (10 ml). Mix thoroughly.  
(sample solutions S9, S10, chromatogram of S9 see Fig. 7).




**Fig. 1** Typical IR spectrum of Tebuconazole




**Fig. 2** Analytical Standard Tebuconazole




**Fig. 3** Technical Material (TC)




**Fig. 4** Wetable powder (WP)



**Fig. 5** Suspension Concentrate (SC)



**Fig. 6** Emulsifiable Concentrate (EC)



**Fig. 7** Emulsion in Water Concentrate (EW)

\