


## EMAMECTIN BENZOATE

XXX



R = CH<sub>2</sub>CH<sub>3</sub> for emamectin B1a benzoate

R = CH<sub>3</sub> for emamectin B1b benzoate

<i>ISO Common Name</i>	Emamectin benzoate
<i>Chemical Name</i>	(4''R)-4''-deoxy-4''-(methylamino) avermectin B1 benzoate
<i>CAS Number</i>	Emamectin benzoate: 155569-91-8 Emamectin B <sub>1a</sub> benzoate: 138511-97-4 Emamectin B <sub>1b</sub> benzoate: 138511-98-5
<i>Empirical formula</i>	Emamectin B <sub>1a</sub> benzoate: C <sub>56</sub> H <sub>81</sub> NO <sub>15</sub> Emamectin B <sub>1b</sub> benzoate: C <sub>55</sub> H <sub>79</sub> NO <sub>15</sub>
<i>Molecular mass</i>	Emamectin B <sub>1a</sub> benzoate: 1008.3 Emamectin B <sub>1b</sub> benzoate: 994.2
<i>m.p.</i>	141-146 °C
<i>v.p</i>	0.004 mPa (21 °C)
<i>Solubility</i>	In water 0.024 g/L at 25 °C; soluble in acetone and methanol
<i>Stability</i>	Stable to hydrolysis at pH 5,6,7 and 8 (25 °C). Photodegrades rapidly.

*Description*                      White to off-white powder

*Formulation*                      -

## EMAMECTIN BENZOATE TECHNICAL

xxx/TC/(M)/-

**1 Sampling.** Take at least 100 g.

### **2 Identity tests**

**2.1 HPLC.** Use the HPLC method below. The relative retention time of the emamectin benzoate in the sample solution should not deviate by more than 1.5% from that of the calibration solution.

**2.2 Infrared.** Prepare potassium bromide discs for the technical sample and emamectin benzoate reference substance. Scan the discs from 4000 to 400  $\text{cm}^{-1}$ . The spectrum from the sample should not differ significantly from that of the reference substance.

### **3 Emamectin Benzoate**

#### **OUTLINE OF METHOD**

Emamectin benzoate is determined by reversed phase high performance liquid chromatography using UV detection at 245 nm and external standardization.

#### **REAGENTS**

*Emamectin benzoate:* reference standard of known purity

*Acetonitrile:* HPLC grade

*Methanol:* HPLC grade

*Ammonium hydroxide:* HPLC grade, 25%~30%

*Ammonia solution:* Ammonium hydroxide : water= 1 : 300 (v/v)

*Water:* HPLC grade

**Calibration solutions.** Weigh in duplicate (to the nearest 0.1 mg) into a volumetric flask (100 ml) about 50 mg of emamectin benzoate standard ( $s$  mg). Dilute to volume with methanol. Mix thoroughly. Filter a portion of each sample solution with a 0.45  $\mu\text{m}$  filter prior to analysis (calibration solutions  $C_A$  and  $C_B$ ).

## APPARATUS

*High performance liquid chromatograph* equipped with a UV detector capable for operation at 245 nm, a constant-temperature column compartment and an injection system capable of injecting 5  $\mu$ l.

*Column stainless steel* 250 x 4.6 mm (i.d.), packed with C<sub>18</sub> 5  $\mu$ m or equivalent with the same selectivity.

*Filtering apparatus* disposable plastic syringes (or equivalent) fitted with 0.45  $\mu$ m filters.

*Electronic integrator or data system*

## PROCEDURE

### (a) Liquid chromatographic conditions (typical):

<i>Column</i>	stainless steel, 250 x 4.6 mm (i.d.), packed with C <sub>18</sub> , 5 $\mu$ m, or equivalent
<i>Mobile phase</i>	methanol : acetonitrile : ammonia solution, 25:55:20(v/v/v)
<i>Column temperature</i>	30°C $\pm$ 2°C
<i>Flow rate</i>	1.2 ml/min
<i>Detector wavelength</i>	245 nm
<i>Injection volume</i>	5 $\mu$ l
<i>Retention time</i>	Emamectin B <sub>1b</sub> : approximately 16.0 min, Emamectin B <sub>1a</sub> : approximately 22.0 min.
<i>Run time</i>	35 min

**(b) System equilibration.** Inject 5  $\mu$ l portion of calibration solution C<sub>A</sub> until the response factors (*fi*) obtained for two consecutive injections differ by less than 1.5%. Then inject 5  $\mu$ l portions of calibration solution C<sub>B</sub>. The response factor (*fi*), for two consecutive injections should not deviate by more than 1.5% from that of solution C<sub>A</sub>, otherwise prepare new calibration solutions.

**(c) Sample preparation.** Prepare solutions in duplicate for each sample. Weigh

(to the nearest 0.1 mg) sufficient sample ( $w$  mg) to contain about 50 mg of emamectin benzoate into a volumetric flask (100 ml). Dilute to the volume with methanol. Mix thoroughly. Filter a portion of each sample solution with a 0.45  $\mu\text{m}$  filter prior to analysis (sample solutions  $S_1$  and  $S_2$ ).

**(d) Determination.** Inject in duplicate 5  $\mu\text{l}$  portions of each sample solution bracketing them by injections of the calibration solutions as follows:

$C_A, S_1, S_1, C_B, S_2, S_2, C_A, \dots$

**(e) Calculation.** Calculate the mean value of each pair of calibration response factors  $f$ , bracketing the two injections of a sample, and use this value for calculating the emamectin benzoate contents of the bracketed sample injections.

$$f_i = \frac{s \times P}{H_s}$$

$$\text{Content of Emamectin benzoate} = \frac{H_w \times f}{W} \text{ (g/kg)}$$

where:

$f_i$  = individual response factor

$f$  = mean response factor

$H_s$  = peak area of emamectin (peak  $B_{1b}+B_{1a}$ ) in the calibration solution

$H_w$  = peak area of emamectin (peak  $B_{1b}+B_{1a}$ ) in the sample solution

$s$  = mass of emamectin benzoate reference standard in the calibration solution (mg)

$w$  = mass of sample taken (mg)

$P$  = purity of the emamectin benzoate reference standard (g/kg)

Calculate the ratio of  $B_{1a}$  to  $B_{1b}$  of emamectin benzoate in the sample:

$$\alpha(B_{1a}/B_{1b}) = \frac{H_{W_{B_{1a}}}}{H_{W_{B_{1b}}}}$$

where:


$\alpha(B_{1a}/B_{1b})$  = ratio of  $B_{1a}$  to  $B_{1b}$  of emamectin benzoate in the sample

$H_{W_{B_{1a}}}$  = peak area of emamectin (peak  $B_{1a}$ ) in the sample solution


$H_{WB_{1b}}$  = peak area of emamectin (peak B<sub>1b</sub>) in the sample solution

**Repeatability r** = XXX-XXX g/kg at an active ingredient content of  
XXX – XXX g/kg


**Reproducibility R** = XXX-XXX g/kg at an active ingredient content of  
XXX – XXX g/kg



**Fig. 1** FTIR spectrum of emamectin benzoate standard



**Fig. 2** HPLC Chromatogram of emamectin benzoate standard



**Fig. 3** HPLC Chromatogram of emamectin benzoate TC