



**Acetonitrile and 3-Picoline in Technical
Chlorantraniliprole and End-Use Formulations
Relevant Impurity Method**

Mary Ellen P. McNally, Ph.D.
July 6th, 2023

DRAFT


ACETONITRILE

<i>Common name</i>	<i>Acetonitrile, methyl cyanide, cyanomethane</i>
<i>Chemical name</i>	Acetonitrile (IUPAC and CAS), CAS Number: 75-05-8
<i>Empirical formula</i>	CH ₃ CN
<i>Molecular Weight</i>	41.053 g/mol
<i>b.p.</i>	80-82°C.
<i>Description</i>	Colorless liquid.
<i>Solubility</i>	<i>Miscible with water</i>
<i>Stability</i>	Stable

3-PICOLINE

<i>Common name</i>	3-Picoline, 3-methyl pyridine
<i>Chemical name</i>	3-Methylpyridine (IUPAC); Pyridine, 3-methyl (CAS) (CAS 108-99-6)
<i>Empirical formula</i>	C ₆ H ₇ N

<i>Molecular Weight</i>	93.13 g/mL
<i>b.p.</i>	144°C
<i>Solubility 20°C</i>	Water soluble
<i>Description</i>	Colorless liquid; Weak base, with a pKa of 5.63

CHLORANTRANILIPROLE

<i>ISO Common name</i>	Chlorantraniliprole
<i>Chemical name</i>	3-bromo-4'-chloro-1-(3-chloro-2-pyridyl)-2'-methyl-6'-(methylcarbamoyl)pyrazole-5-carboxanilide (IUPAC) 3-Bromo-N-[4-chloro-2-methyl-6-[(methylamino)carbonyl]phenyl]-1-(3-chloro-2-pyridinyl)-1H-pyrazole-5-carboxamide (CAS 50000845-7)
<i>Empirical formula</i>	C ₁₈ H ₁₄ BrCl ₂ N ₅ O ₂
<i>RMM</i>	483.15
<i>m.p.</i>	208-210°C
<i>Solubility in various solvents at 20°C</i>	1.0 µg/mL in water 124 mg/mL in dimethylformamide 3.4 mg/mL in acetone

2.5 mg/mL in dichloromethane
1.7 mg/mL in methanol
1.1 mg/mL in ethyl acetate
0.71 mg/mL in acetonitrile
0.39 mg/mL in n-octanol
0.16 mg/mL in o-xylene
<0.1 µg/mL in n-hexane

Description Fine crystalline off-white powder

Stability Aqueous hydrolysis: Chlorantraniliprole is stable at pH 4 and 7; at pH 9, chlorantraniliprole hydrolyzed with a half-life of approximately 10 days.

Formulations
600 SC, Flowable Concentrate for Seed Treatment
200 SC, Suspension Concentrate

1. **SAMPLING.**

As received.


2. **IDENTITY TESTS.**

1. **GC:** Using the GC method described below. The relative retention time of the 3-picoline and acetonitrile peaks in the sample solution should not deviate by more than 5% from that of the standard solution. The mass spectrum measured from these peaks should match that obtained from the standard substance as shown in the example mass spectra below.

**Example Mass spectrum of Acetonitrile and 3-Picoline in 200 SC Formulated
Sample of Chlorantraniliprole.**

A. Acetonitrile


Sample ID: 90907-SP-Cora



B. 3-Picoline

DRAFT

Sample ID: 90907-SP-Cora



2. Infrared (IR).

Attenuated Total Reflectance (ATR) was used to determine the FT-IR spectra of the standard material of acetonitrile and 3-picoline. The ATR diamond is cleaned with a wipe or cotton swab immersed in water, then either ethanol, methanol, or isopropanol, followed by acetone. The background is collected before each sample, then with a pipette the sample is transferred directly to cover the diamond ATR crystal.

The sample is scanned using the following conditions and saved:

Range: 4000-650 cm^{-1}


Scan Number: 32

Units: Absorbance


Resolution: 4.00 cm^{-1}

The expected levels of acetonitrile and 3-picoline are not visible in the technical material or formulated products of chlorantraniliprole by ATR.

Infrared Spectrum of Acetonitrile Standard



Infrared Spectrum of 3-Picoline Standard



3. ACETONITRILE AND 3-PICOLINE IN CHLORANTRANILIPROLE

OUTLINE OF GC METHOD

A solution of the sample in *N,N*-dimethylacetamide (DMAC) is separated by capillary gas chromatography. Acetonitrile and 3-Picoline are detected with a flame ionization detector (FID). The weight percent of acetonitrile and 3-picoline in the chlorantraniliprole technical and formulated samples are determined by comparing the peak area with a calibration curve (peak area vs. concentration) prepared from the analysis of standard solutions.

The method is applicable to the determination of acetonitrile and 3-picoline in chlorantraniliprole technical samples and end-use products. The applicable concentration range for acetonitrile is approximately 0.004% - 1.65% by weight and for 3-picoline is 0.004% - 1.64% by weight for a 35.0 mg/mL sample solution of technical material. The range can be extended by modifying the standard or sample

concentrations.

REAGENTS

Acetonitrile, HPLC-GC Grade, Purity >99.99%

3-Picoline, Analytical Standard, Purity >99.50%

N,N-Dimethylacetamide, HPLC-GC Grade

Calibration solutions.

Seven standards are prepared to demonstrate linearity. For routine analyses once an acceptable 7-point calibration is produced, it is acceptable to use a three-point calibration that covers the range of analyte concentration in the samples to be analyzed.

Weigh 60 mg (\pm 10 mg) of acetonitrile standard solvent into a 100mL volumetric flask prefilled with 40 mL of *N,N*-Dimethylacetamide (DMAC). Tare the flask and weigh 55 mg (\pm 10 mg) of 3-picoline. Dilute to volume with DMAC. This will be the stock standard. The nominal concentrations for acetonitrile and 3-picoline in this solution will be approximately 0.60 and 0.55 mg/mL, respectively.

Make the following dilutions into separate volumetric flasks. Dilute to volume with DMAC.

The volumetric glassware used should be type “Class A” glassware.

	Preparation	ACN concentration, mg/mL	3-Picoline concentration, mg/mL
Standard 1	Standard Stock Solution	0.600	0.550
Standard 2	25 mL Stock into 50 mL	0.300	0.275
Standard 3	15 mL Stock into 50 mL	0.180	0.165
Standard 4	10 mL Stock into 50 mL	0.120	0.110
Standard 5	5 mL Stock into 50 mL	0.060	0.055
Standard 6	3.5 mL Stock into 50 mL	0.042	0.039
Standard 7	2 mL Stock into 50 mL	0.024	0.022

Preparation of Calibration Curve(s)

Equilibrate the instrument and column for one hour at the initial temperature. Cycle the oven one time through the temperature program prior to analysis. After the baseline has stabilized, inject an assay and impurity analytical standard solution to verify the retention times of each component.

Inject 1 μL of each standard and sample solution. For quantitative determinations, duplicate injections of standards and duplicate weighing of samples and duplicate injections of each weighing are recommended.

Prepare a calibration curve for each component by plotting the average peak area vs. concentration (mg/mL), corrected for purity, for each standard solution.

Using the method of least squares, calculate the equation for the best line fit for each component. Typically, the correlation coefficient of a least squares equation is 0.999 or better. A computer data system, with least squares reduction capability, is recommended to generate calibration curves (i.e., $y = mx + b$) and the statistical analyses of the curves (i.e., correlation coefficients).

APPARATUS

Gas chromatograph: Gas chromatograph capable of capillary column installation and equipped with a split/splitless inlet, an automated sample injector capable of injecting a 1 μL aliquot, constant temperature oven with temperature programming capability, flame ionization detector (FID), and integrator/recorder or some other data handling system.

GC Column: Agilent DB Wax (30 m \times 0.250 mm) \times 0.50 μm film thickness (P/N: 122-7033)

Ultrasonic bath

Filters PTFE: 0.20 μm

Class A Volumetric flasks with stoppers

Class A Analytical pipets

Analytical balance: capable of measuring ± 0.1 mg.

PROCEDURE

(a) Operating Conditions

Oven

Initial Temp:	60° C
Initial Hold Time:	3.0 minutes

Program Rate 1:	15.0° C / minute
Final Temperature 1:	160° C
Intermediate Hold Time:	0.0 minute
Program Rate 2:	25° C / minute
Final Temperature 2:	240° C
Final Hold Time:	3.0 minutes
Total Run Time:	15.9 minutes

Inlet Conditions

Mode:	Split
Inlet Temperature:	150° C
Pressure:	19.9 psi
Split Ratio:	10:1
Carrier Gas Type:	Hydrogen
Injection Volume:	1.0 µL

The boiling points of acetonitrile and 3-picoline are 81-82 and 144°C, respectively. The injector temperature is set at 150°C to ensure volatilization and to minimize interferences.

The split flow determines the percentage of sample loaded onto the column. The split ratio is the ratio of split flow to column flow. A ratio of 20 is required to obtain optimum detector sensitivity from the FID detector. Adjust this as needed to optimize the instrument, since instrument performance varies.

Column Conditions

Column:	Agilent DB Wax (30 m × 0.250 mm) × 0.50 µm film thickness
Mode:	Constant Pressure
Initial Flow:	5.0 mL/min
Nominal Initial Pressure:	19.9 psi

Flame Ionization Detector Conditions

Temperature:	240° C
Hydrogen Flow:	33 mL/min
Air Flow:	300 mL/min
Mode:	Constant makeup flow

Makeup Flow: 30 mL/minute
 Makeup Gas: Nitrogen
 Approximate Retention Times: Acetonitrile – 4.1 minutes
 3-Picoline – 8.4 minutes
 Peakwidth* (response time): > 0.03 min (0.5 sec)

*The peakwidth parameter may be instrument-specific. Ensure a data collection rate that permits at least 50 collected points across each peak.

Analysis Stop Time 16.0 minutes

(b) Sample preparation.

Before sub-sampling liquid samples, mix thoroughly by swirling for approximately 1 to 2 minutes.

For solid samples, grind and mix all samples thoroughly before weighing the test portions for analysis.

Refer to the table below and accurately weigh (to at least 0.1 mg precision) the amount of sample indicated into a 10 or 50 mL volumetric flask.

<u>Sample</u>	<u>Nominal Amount</u>	<u>Volume</u>
Chlorantraniliprole Technical	350 ± 10 mg	10 mL
Chlorantraniliprole 200 g/L Suspension Concentrate	5000 ± 50 mg	50 mL
Chlorantraniliprole 600 g/L Suspension Concentrate	1750 ± 20 mg	50 mL

Dilute each flask to volume with DMAC and ultrasonicate for ten minutes.

Not all formulation inert ingredients will completely dissolve.. Swirling during the final 5 minutes of ultrasonication aids the dissolution

Allow the solution(s) to equilibrate to room temperature and syringe filter the sample into a GC sample vial and inject into the GC.

Determination. Equilibrate the column until a stable baseline is obtained. Inject, in duplicate, 1 µL each of a solvent blank (DMAC), standards, and samples, bracketing the samples with standard solutions.

Calculation. Determine the analyte (i.e., acetonitrile and/or 3-picoline) concentration (mg/mL) from the peak area of the analyte using the equation from respective calibration curve, $y = mx + b$.

$$\text{Analyte concentration (x)} = \frac{\text{Analyte peak area (y) - y-intercept(b)}}{\text{Slope(m)}}$$

DRAFT


The analyte weight % can be calculated as follows:

$$\text{Weight \% Analyte} = \frac{(\text{mg/mL Analyte})(\text{Final Volume (mL)})}{\text{Sample weight (mg)}} \times 100$$


DRAFT

EXAMPLE CHROMATOGRAMS:


GC Chromatogram of a Blank DMAC Solution



GC Chromatogram of Standard Solution 4 Containing 0.043 mg/mL of Acetonitrile and 0.045 mg/mL of 3-Picoline in DMAC



GC Chromatogram of Chlorantraniliprole Technical Sample containing 35.12 mg/mL (351.19 mg/10mL) technical a.i. in DMAC



GC Chromatogram of Chlorantraniliprole 600 SC Formulation Sample containing 35.05 mg/mL (1752.6 mg/50mL) technical a.i. in DMAC

