

# DETERMINATION OF RESIDUES OF TRANSFLUTHRIN AND PERMETHRIN IN AIR WHILE USING AN AEROSOL FORMULATION

BY

DR.A.RAMESH Ph.D.,D.Sc.,C.Chem.,MRSC.,

&

MR. GUNDOJU SURESH M.Sc., CAQM.,

HEAD, DEPARTMENT OF ANALYTICAL CHEMISTRY  
INTERNATIONAL INSTITUTE OF BIOTECHNOLOGY AND TOXICOLOGY  
INDIA

# OBJECTIVES

- To develop an analytical method to quantify the residues of Permethrin and Transfluthrin in air samples
- To establish an effective procedure to collect air samples following the use of Transfluthrin 0.6% w/w + Permethrin 0.8% w/w Aerosol in a closed chamber

# AEROSOL DISPENSER SYSTEM



- Dispenser works on 15 minutes interval
- Declared volume in the can is 154 grams
- Life of the can is 28.9 ~ 30 days
- Spray Rate is 0.055 gram/spray




# STUDY DESCRIPTION

- This residue study was conducted using an automatic aerosol dispenser
- The sprayer was switched on for 24 hours at a spray rate of 0.055g/spray once every 15 minutes
- The residues of permethrin and transfluthirun in air were estimated for a period of 24hrs in a pre-determined intervals (0, 2, 4, 6, 10 and 24 hours)
- Air Samples were collected at 1.5 m height from the floor using cartridge

# SCHEMATIC DIAGRAM - AEROSOL SPRAY

Height – 2.1 m  
Floor area – 13 m<sup>2</sup>  
Total Volume – 27.3 m<sup>3</sup>



- 500 mL of air sample was withdrawn from the chamber approx. 3-5 minutes after the spray
- The collected air sample was pre-concentrated and analyzed by a validated GC-ECD method

# TEST ITEM DETAILS

1. Name of the compounds:
  - i. Transfluthrin
  - ii. Permethrin
  
2. Nature of the compound : Insecticide
  
3. Type of formulation : Aerosol
  
4. Percentage of active ingredient :
  - i. Transfluthrin – 0.6%
  - ii. Permethrin – 0.8%

# INSTRUMENTS/EQUIPMENTS

- Shimadzu GC-17A GC-ECD, Auto injector AOC - 20i with GC-solution software. M/s Shimadzu Corop., Japan
- Mettler Toledo analytical balance AG-245, capable of weighing 0.01 mg supplied by M/s.Mettler Toledo, Switzerland.
- Buchi temperature controlled vacuum rotary evaporator supplied by M/s.Buchi Rotovapour, Switzerland.

# GAS CHROMATOGRAPHY ANALYSIS PARAMETERS

Detector -<sup>63</sup>Ni Electron Capture Detector (ECD)

Column used -DB-210 Megabore (30m length x 0.53 mm I.D., 1.0 μm film thickness).

## Temperature conditions

Oven -150°C

Injector -240°C

Detector -260°C

## Gas flow rate

Nitrogen -10 mL/min

Makeup -30 mL/min

## Column temperature program

Rate	Temperature °C	Hold time (min)
-	150	5
20	220	6.5

## Retention time (approximate)


Transfluthrin (Standard) -6.1 minutes

Permethrin (Standard) Cis -10.5 minutes

Trans -11.2 minutes




# REPRESENTATIVE CHROMATOGRAM OF GC-ECD



# METHOD VALIDATION - LINEARITY

## CALIBRATION CURVE-TRANSFLUTHRIN+PERMETHRIN



Calibration range 0.001 – 2 mg/L  
LOQ = 0.001mg/L

# METHOD VALIDATION - RECOVERY

Spikings: 1, 0.1 and 0.001 mg/L level on Carbon cartridge

Suction of 500mL of Clean air

Eluted with 10mL of Acetone

Analyzed 1 $\mu$ L by GC-ECD

# METHOD VALIDATION - RECOVERY


Fortified concentration (mg/L)	Transfluthrin Recovery (%)	Permethrin Recovery (%)	Transfluthrin mean Recovery (%) $\pm$ SD	Permethrin mean recovery (%) $\pm$ SD
0.001	89.98	96.88	93.46 $\pm$ 3.34	93.50 $\pm$ 4.04
0.001	93.75	89.03		
0.001	96.65	94.59		
0.1	90.5	92.38	92.51 $\pm$ 2.49	94.67 $\pm$ 2.00
0.1	95.29	95.62		
0.1	91.74	96.02		
1	94.48	93.08	94.36 $\pm$ 2.26	94.38 $\pm$ 1.71
1	92.05	96.32		
1	96.56	93.75		

SD - Standard deviation


# METHODOLOGY – SAMPLE COLLECTION

- An automatic insect control aerosol system was fixed on wall in one corner of the testing chamber.
- Air samples were collected by placing the carbon cartridge at **5 centimeters** distance from the sprayer nozzle.
- 500 ml of air samples were withdrawn from the chamber at **0-10, 10-20, 30-40 and 60-70 and 90-100 seconds** after the spray at a rate of 3 L /min for 10 seconds using a cartridge connected to a vacuum pump equipped with flow regulator
- Trapped residues were eluted using 10 ml acetone twice.
- Evaluated the residues of transfluthrin and permethrin using a validated **GC-ECD** method.

# AIR CONCENTRATION CHAMBER - SETUP



# STUDY DESIGN – AIR CONCENTRATION RECOVERY TRAP



# PRELIMINARY INVESTIGATIONS

- No residues were found when carbon trap was kept at 20, 30, 45 and 60 cm away from cartridge
- Non-recovery of residues at longer distance can be attributed due to:
  - Low concentration of residues in air samples
  - Requirement of huge volume air sample to concentrate the residues on the cartridge in a shortest plausible time
  - 5 cm distance between tip of the nozzle and the carbon trap was found optimum.




**PRE CONCENTRATION OF  
TRANSFLUTHRIN AND PERMETHRIN  
FOLLOWED BY 0.055 G/SPRAY USING CARBON CARTRIDGE  
KEPT AT VARIED DISTANCES**

Time gap between Spray and Sample Collection (sec)	Sample Collection Duration (sec)	Sample Collection flow rate (L/min)	Volume of air (mL)	Transfluthrin active content (mg/L)	Permethrin active content (mg/L)
0-10	10	3	500	0.1559	0.1985
10-20	10	3	500	0.1228	0.1824
30-40	10	3	500	0.0864	0.1121
60-70	10	3	500	0.0254	0.0416
90-100	10	3	500	BLQ	BLQ

BLQ - Below Limit of Quantification


# Air Concentrations of Actives in 1 m3 Chamber



# RESULTS WITH SINGLE SPRAY

- The residues of permethrin and transfluthrin were found to be below detectable level in all sampling occasions
- Non-recovery of residues can be attributed due to:
  - Low concentration of the residues in air
  - The actives may reduces quit quickly after the spray
  - Huge volume of air sample may be required to pre-concentrate the residues on cartridge
  - Sampling and extraction method need to be optimized
- The present study is designed to minimize such influential parameters Conclusion

# RESIDUE ACCUMULATION IN A ROOM DUE TO 8 HRS SPRAY



# EXPERIMENTAL DETAILS

- ✓ Aerosol sprayer was switched on for 8 hours in a dark 1m<sup>3</sup> chamber by fixing the device on wall in one corner.
- ✓ Spray Rate is 0.55g/spray once in 15 minutes time intervals
- ✓ Air sampling was done from different locations of the chamber - Top, Middle & Bottom.

Insecticide formulation	Time of collection (hours)
Permethrin 0.8% w/w + Transfluthrin 0.6% w/w	0.5, 1, 2, 4, 6 and 8

# EXPERIMENTAL DETAILS

- ✓ 500 CC Volume of air samples were collected using cartridge in each location from the room
- ✓ Collected samples were dissolved in 10mL of acetone and quantified by GC-ECD.
- ✓ Once again all the samples were pooled and reduced the volume to 1mL and analyzed by GC-ECD to determine the total residue concentration.
- ✓ Surface concentration was measured after completion of 8 hours
- ✓ Floor of the chamber was covered with poly propylene sheet
- ✓ Poly propylene sheet was cut into pieces and extracted with 250 mL of acetone using end-over-end mechanical shaker.
- ✓ Volume of acetone was reduced to 1 mL using rotary evaporator and analyzed by GC-ECD

# CONCENTRATION OF TRANSFLUTHRIN AND PERMETHRIN IN AIR AT DIFFERENT LOCATIONS IN THE ROOM

Sampling Occasion (hours)	Locations					
	Transfluthrin (mg/L)			Permethrin (mg/L)		
	Top	Middle	Bottom	Top	Middle	Bottom
0.5*	BLQ	BLQ	BLQ	BLQ	BLQ	BLQ
1*	BLQ	BLQ	BLQ	BLQ	BLQ	BLQ
2	BLQ	BLQ	0.0013	BLQ	BLQ	0.0016
4	0.0011	0.0017	0.0019	0.0013	0.0015	0.0021
6	0.0015	0.0024	0.0036	0.0019	0.0028	0.0047
8	0.0028	0.0062	0.0081	0.0033	0.0071	0.0093

BLQ - Below Limit of Quantification

\* - Pooled samples were analysed

# DEPOSITIONS OF TRANSFLUTHRIN AND PERMETHRIN ON FLOOR IN THE CHAMBER

Sampling Occasion (hours)	Residue (mg/L)	
	Transfluthrin	Permethrin
8	0.282	0.683



# CONFIRMATION OF RESIDUES BY GC – MS

Instrument - Shimadzu GCMS QP5050A  
Column - DB-5 (30 m x 0.25 mm I.D x1.0 µm)

## Temperature conditions

Oven -150°C  
Injector -240°C  
Detector -260°C

## Column temperature program

Rate	Temperature °C	Hold time (min)
-	150	5
20	220	6.5

## Gas flow rate

Helium -1.0  
Split ratio - 1:20

## Ion Monitored

Transfluthrin - 127,163 m/z  
Permethrin - 163,183 m/z

## Retention time (approximate)

Transfluthrin (Standard) -6.1 minutes  
Permethrin (Standard) Cis -14.5 minutes  
Trans -15.2 minutes



# MS SPECTRA - PERMETHRIN

